



Report Test Resistivity Test

Identification	Identification	ISO 1853 - 20 mm Electrometer
Job : (Generic) Order : order A Lot : lot A Batch : 1 Date : 26/10/2009	Product : Compound 01 Cod. product : 001 Treatment : Original State Customer : <Generic> Cod.Customer : 0 Operator : Administrator	Distance of Electrodes: 20.00 mm Sample width: 9.87 mm Sample thickness: 2.04 mm Test Time: 60.00 sec

Notes :

Test N°	Wt mm2	I mA	V V	Rv Ohm*m	Rv Ohm*cm
Max Tol	0.00	0.000e+0	0.000e+0	0.000e+0	0.000e+0
Min Tol	0.00	0.000e+0	0.000e+0	0.000e+0	0.000e+0
1	20.21	2.788e+1	8.100e-1	2.936e-2	2.936e+0
2	20.13	2.785e+1	8.099e-1	2.927e-2	2.927e+0
3	20.13	2.784e+1	8.324e-1	3.009e-2	3.009e+0
Max	20.21	2.788e+1	8.324e-1	3.009e-2	3.009e+0
Min	20.13	2.784e+1	8.099e-1	2.927e-2	2.927e+0
Mean	20.16	2.786e+1	8.174e-1	2.957e-2	2.957e+0
Dev.Std	0.0462	0.0208	0.0130	0.0004	0.0450
Cp					
Cpk					

Signature _____

