

# CERTIFIED SAMPLES

Gibitre Instruments는 쇼어 및 IRHD 경도 검증을 위한 인증 샘플을 제공합니다.



### 개요

인증된 샘플을 사용하면 교정 사이의 기간에 경도 시험기 판독 값의 일치 여부를 정기적으로 확인할 수 있습니다.

주기적 검증은 테스트 파라미터로서의 경도의 중요성을 고려하여 유용합니다.

### 제품의 특성

엘라스토머 제품의 경도는 온도에 크게 영향을

받습니다. 이러한 이유로 Gibitre가 생산 한 샘플은 손의 체온이 테스트 영역으로 전달하지 않고 쉽게 처리 할 수 있는 모양을 가지고 있습니다.

샘플에는 측정에 사용되는 공인된 경도 시험기에 교정 인증서가 함께 제공됩니다.

시료에는 온도 변화 및 빛으로부터 시료를 보호할 수 있는 절연 보호 케이스가 제공됩니다.

사용 가능한 구성

Gibitre는 Shore A, Shore D, IRHD-N 및

IRHD-Micro 경도를 테스트 하기 위한 샘플을 제공합니다.

다음 구성을 사용할 수 있습니다.

- 선택한 경도 범위 (약 40 - 50 - 60 - 80 - 90) 내에서 경도가 다른 5 개의 샘플을 포함한 상자
- 사용 가능한 경도 중 하나와 함께 하나의 단일 샘플이 들어있는 상자

**Shape of the samples:** The shape of the samples has been developed to permit easy handling without heat transmission to the test area

**Protection Box:** The wooden box ensures protection against light and temperature variations

**Sample identification:** The samples have unique identification code to

permit the traceability of the calibration

**Calibration Report:** The calibration report is issued by Gibitre Instruments and includes the traceability to the officially-calibrated hardness tester used for the measures

